



RX Self-expanding Peripheral Stent

Patient Implant Card

INSTRUCTIONS

Please carry this card at all times. If you receive medical care, show your **Patient Implant Card** to any medical personnel who may treat you.

If you have questions regarding your Misago, contact the implanting physician or Terumo Medical Corporation at 1-800-888-3786 or visiting www.terumomedical.com.



Non-clinical testing has demonstrated Misago is MR Conditional. A patient with this device can be safely scanned in an MR system meeting the following conditions:

- Static magnetic field of 1.5 T or 3 T
 - Maximum spatial field gradient of 4,000 gauss/cm (40 T/m)
 - Maximum MR system reported, whole body averaged specific absorption rate (SAR) of 2 W/kg for Normal Operating Mode
- Under the scan conditions defined above, Misago is expected to produce a maximum temperature rise of 3.2°C after 15 minutes of continuous scanning.

Artifact Information

In non-clinical testing, the image artifact caused by the device extends approximately 2 mm from Misago when imaged with a gradient echo pulse sequence and a 3.0T MRI system.

Notify your doctor before you have an MRI scan. If possible, exclude the body area from radiofrequency (RF) exposure where the stent is implanted. MR image quality may be compromised if the area of interest is in the exact same area or relatively close to the position of the stent. The lumen of Misago can be visualized on T1-weighted spin echo and gradient echo pulse sequences.

Recommended Registration Information

Please consider registering your stent implant information with the Medic Alert Foundation by calling 1-888-633-4298 or visiting www.medicalert.org.

REV Dec 2016

Patient Name

Date of Stent Implant

Stent Location(s)

Implanting Physician

Hospital

Address

Telephone

Primary Physician

Telephone

TERUMO is a trademark of TERUMO CORPORATION. ©: Registered Trademark.